

PARAMETRI CONTROLLI EMISSIONI DA SORGENTE FISSA

DETERMINAZIONE	METODO	ACC
Acido cloridrico	UNI EN 1911:2010 Met C escluso campionamento	√
Acido Fluoridrico	PrEN 17340:2019 met. C.4 escluso campionamento	√
Ammoniaca	UNI EN ISO 21877:2020 Annex C escluso campionamento	√
Anidride solforosa	UNI EN 14791:2017 Met A escluso campionamento	√
PolicloroDibenzodiossine (PCDD) e PolicloroDibenzofurani (PCDF): 2,3,7,8 Tetra-CDD; 1,2,3,7,8 Penta-CDD; 1,2,3,4,7,8 Esa-CDD; 1,2,3,6,7,8 Esa-CDD; 1,2,3,7,8,9 Esa-CDD; 1,2,3,4,6,7,8, Epta-CDD; Octa-CDD; 2,3,7,8 Tetra-CDF; 1,2,3,7,8 Penta-CDF; 2,3,4,7,8 Penta-CDF; 1,2,3,4,7,8 Esa-CDF; 1,2,3,6,7,8 Esa-CDF; 1,2,3,7,8,9 Esa-CDF; 2,3,4,6,7,8 Esa-CDF; 1,2,3,4,6,7,8 Epta-CDF; 1,2,3,4,7,8,9 Epta-CDF; Octa-CDF	UNI EN 1948-2:2006 + UNI EN 1948-3:2006	√
Formaldeide, acetaldeide, acroleina e benzaldeide	EPA 8315A rev.1 1996	√
Benzene, Toluene, etilbenzene, o-xilene, m-xilene, p-xilene	UNI CEN/TS 13649:2014	√
PCB: PCB 28, PCB 52, PCB101, PCB138, PCB153, PCB180, PCB105, PCB118(5000-50000 pg); PCB77, PCB114, PCB123, PCB189 (200-50000 pg); PCB81(100-50000 pg); PCB126, PCB169 (20-50000 pg).	UNI EN 1948-4:2010	√
Idrocarburi policiclici aromatici (IPA): Antracene, Benzo(a)antracene, Benzo(a)pirene, Benzo(b+j)fluorantene, Benzo(e)pirene, Benzo(ghi)perilene, Benzo(k)fluorantene, Crisene, Dibenzo(a,h)antracene, Fenantrene, Fluorantene, Indeno(1,2,3-cd)pirene, Perilene, Pirene.	ISO 11338-2:2003	√

LEGENDA

ACC

prova accreditata ISO 17025 se spuntata